

DC Electronic Load

PEL-3000 Series

QUICK START GUIDE

GW INSTEK PART NO. 82EL-31110MA1



ISO-9001 CERTIFIED MANUFACTURER

GW INSTEK

This manual contains proprietary information, which is protected by copyright. All rights are reserved. No part of this manual may be photocopied, reproduced or translated to another language without prior written consent of Good Will Corporation.

The information in this manual was correct at the time of printing. However, Good Will continues to improve its products and therefore reserves the right to change the specifications, equipment, and maintenance procedures at any time without notice.

Good Will Instrument Co., Ltd. No. 7-1, Jhongsing Rd., Tucheng Dist., New Taipei City 236, Taiwan.

S SAFETY INSTRUCTIONS

This section contains the basic safety symbols that may appear on the accompanying User Manual CD or on the instrument. For detailed safety instructions and precautions, please see the Safety Instructions chapter in the user manual CD.

Safety Symbols

These safety symbols may appear in the user manual or on the instrument.



Warning

Warning: Identifies conditions or practices that could result in injury or loss of life.



Caution

Caution: Identifies conditions or practices that could result in damage to the instrument or to other properties.



DANGER High Voltage



Attention Refer to the Manual



Protective Conductor Terminal



Earth (ground) Terminal



Do not dispose electronic equipment as unsorted municipal waste. Please use a separate collection facility or contact the supplier from which this instrument was purchased.

Power Cord for the United Kingdom

When using the instrument in the United Kingdom, make sure the power cord meets the following safety instructions.

NOTE: This lead/appliance must only be wired by competent persons.



WARNING: THIS APPLIANCE MUST BE EARTHED
IMPORTANT: The wires in this lead are coloured in accordance with the following code:

Green/ Yellow: Earth

Blue: Neutral

Brown: Live (Phase)



As the colours of the wires in main leads may not correspond with the coloured marking identified in your plug/appliance, proceed as follows:

The wire which is coloured Green & Yellow must be connected to the Earth terminal marked with either the letter E, the earth symbol  or coloured Green/Green & Yellow.

The wire which is coloured Blue must be connected to the terminal which is marked with the letter N or coloured Blue or Black.

The wire which is coloured Brown must be connected to the terminal marked with the letter L or P or coloured Brown or Red.

If in doubt, consult the instructions provided with the equipment or contact the supplier.

This cable/appliance should be protected by a suitably rated and approved HBC mains fuse: refer to the rating information on the equipment and/or user instructions for details. As a guide, a cable of 0.75mm² should be protected by a 3A or 5A fuse. Larger conductors would normally require 13A types, depending on the connection method used.

Any exposed wiring from a cable, plug or connection that is engaged in a live socket is extremely hazardous. If a cable or plug is deemed hazardous, turn off the mains power and remove the cable, any fuses and fuse assemblies. All hazardous wiring must be immediately destroyed and replaced in accordance to the above standard.

GETTING STARTED

The Getting Started chapter introduces the instrument's main features, appearance, and set up procedure.

Overview

The PEL-3000 Series is a family of high performance DC electronic loads positioned to test a wide range of different power sources. The DC electronic loads are fully programmable to simulate anything from basic static loads to complex dynamic loads. With the ability to operate independently or in parallel, the PEL-3000 Series is extremely robust and capable of molding to any test environment.

Model Line Up

Model	Operating Voltage (DC)	Current	Power
PEL-3021	1.5V~150V	35A	175W
PEL-3041	1.5V~150V	70A	350W
PEL-3111	1.5V~150V	210A	1050W
PEL-3211 (Booster)	1.5V~150V	420A	2100W

Main Features

Performance	• High slew rates of up to 16A/μS(PEL-3111) for a fast response speed
-------------	--

	• High capacity when used in parallel: 5250W, 1050A (PEL-3111 x 5)/ 9450W, 1890A (PEL-3111 + PEL-3211 x 4) • High resolution – 16 bit
Features	• 7 operating modes: CC, CV, CR, CP, CC+CV, CR+CV, CP+CV • Independent and parallel operation • Fully programmable with normal and fast sequences • Soft start • Dynamic mode • OCP, OVP and other protection features • Remote sense • Integrated meter • Rack-mountable • Load booster
Interface	• USB, RS232 and GPIB • External voltage or resistance control • Front panel trigger out BNC • Front panel current monitoring BNC • Analog external control

Package Contents and Accessories

Standard Accessories

Item	Part Number
User /Programming Manual CD	82EL-31110EA1
Quick Start Guide (this document)	82EL-31110MA1
Load input terminal Cover	PEL-011
Terminal fittings: 2 sets of bolts/nuts/springs/washers (type: M8)	PEL-012
Power Cord x1	Region Dependent

Optional Accessories


Item	Part Number
Rack mount bracket for booster PEL-3211/EIA	GRA-413-E
Rack mount bracket for booster PEL-3211/JIS	GRA-413-J
Rack mount frame for PEL-3021, PEL-3041, PEL-3111/EIA	GRA-414-E
Rack mount frame for PEL-3021, PEL-3041, PEL-3111/JIS	GRA-414-J
300mm frame link cable (for vertically stacked units)	GTL-255
GPIO cable, 2.0m	GTL-248
USB cable. Type A - Type B	GTL-246
Dust filter	PEL-010

Options


Item	Part Number
GPIO Interface	PEL-004

Front Panel

PEL-3021 and PEL-3041



PEL-3111



PEL-3211




Description


- | | |
|---|--|
| 1. Air inlet | 2. LCD Display |
| 3. Function keys | 4. Power key |
| 5. Main/Local key | 6. FUNC/File key |
| 7. Help/Utility key | 8. Short key |
| 9. Load On/Off | 10. Scroll wheel |
| 11. Number pad,
Clear/Lock and Enter
keys | 12. USB port, Preset and
Shift keys |
| 13. I MON OUT, TRIG
OUT | 14. Input terminals |

Rear Panel


PEL-3021 and PEL-3041



PEL-3111




PEL-3211



Description

- | | |
|--------------------------------|------------------------|
| 1. Frame control ports, J1, J2 | 2. Remote sense inputs |
| 3. Rear panel inputs | 4. Exhaust fan |
| 5. Power socket and switch | 6. GPIB (optional) |
| 7. USB device port | 8. USB port |
| 9. RS232C port | |

Display Overview



Description


- | | |
|---------------------------|----------------------------|
| 1. Date and time | 2. Main frame status panel |
| 3. Operation status panel | 4. Soft keys |
| 5. Setting area | 6. Measurement area |

First Time Use Instructions


Use the following procedures when first using the PEL-3000 to power up the instrument, set the internal clock, restore the factory default settings and check the firmware version. Lastly, the Conventions section will introduce you to the basic operating conventions used throughout the user manual.

Power Up

1. Insert the AC power cord into the power socket.
2. Turn the power switch on from the rear panel.
(O → —)



3. If the unit doesn't turn on, press the ON/STBY key on the front panel.
 - The ON/STBY key will go from standby (red) to ON (green).



4. The unit will show the splash screen and then load the settings from when the unit was last powered down.

Load Default Settings

When first using the PEL-3000, recall the factory default settings to ensure the unit is in a known state. See the user manual for a list of the default settings.

1. Press  +  .
File
2. Select *Media/Default* [F1].
3. Select *Factory Default* [F2].
4. Press *Factory Default* [F2] again to confirm.

Setting the Date and Time

The date and time settings are used to time-stamp files when saving files.

1. Press  +  > *Time Set[F4]* to set the date and time.

Settings: Month, Day, Year, Hour, Minute

Updating the Firmware

The PEL-3000 allows the firmware to be updated by end-users. Before using the PEL-3000, please check the GW Instek website or ask your local distributor for the latest firmware. Before updating the firmware, please check the firmware version.

View Firmware Version

1. Press  +  .
2. Select *System/Info[F1]*.
3. The system information is listed in the display.
 - MainFrame Ver: Mainframe firmware version.
 - FPGA Ver: FPGA firmware version
 - PEL-3XXX SN: Serial number of the unit.
 - SCPI Ver: SCPI-compatible version.

Firmware update

1. Press **Shift** + **FUNC** (File).
2. Select *USB* with the *Media [F1]* soft-key.
3. Press the *File Utility [F5]* soft-key.
4. Select the *.UPG upgrade file and press *Select[F1]* twice. Once to select the file and once to confirm.
5. Wait for the update to complete and reset the power.



Warning


Do not turn the load generator off or remove the USB memory when the firmware is being read or upgraded.

Conventions

The following conventions are used throughout the user manual. Read the conventions below for a basic grasp of how to operate the PEL-3000 menu system using the front panel keys.

Soft-menu keys

The F1 to F5 function keys at the bottom of the display correspond directly to the soft-menu keys on top.



Select Sub Menu



Pressing this type of soft-menu key will enter a submenu.

Toggle Parameter or State



This type of soft-menu icon has the function/item on the top of the label and the selected setting or mode on the bottom of the label.

Repeatedly press the associated function key (F1~F5) to cycle through each setting.


For some parameters, a popup window will also appear. Selection of the setting is the same.

Repeatedly pressing the relevant function key (F1~F5) will cycle through each setting.


Parameter Input

The scroll wheel, Enter key and number pad can be used to edit parameter values.

1. Use the scroll wheel to move the cursor to the desired parameter.
 - A scroll bar is shown when there are additional parameters off-screen.
-



2. Press the **Enter** key to select the parameter.
3. Then use the number pad* or scroll wheel** to edit the parameter value.




Clearing a Value

*When editing a parameter with the number pad, pressing the **Clear** key will restore the parameter to the previous value.

Coarse/Fine Adjustment

**When a parameter is highlighted (step 3 above) pressing the scroll wheel will toggle the scroll wheel resolution between fine and coarse.




Note: There is a second method of fine adjustment

that allows you to edit parameters one digit value at a time using the scroll wheel. This is called Cursor mode. Please see the user manual for more information.

Entering Alphanumeric Characters

When renaming files, creating memos or notes, you will be required to enter alphanumeric characters when the character entry screen appears.

- Only alphanumeric characters as well as space [], underscore [_] and minus [-] characters are allowed.
1. Use the scroll wheel to move the cursor to the desired character.



2. Press the  key or *Enter Character*[F1] to select a character.
3. To delete a character, press *Back Space*[F2].
4. To save the file name or memo, press *Save*[F3].

SPECIFICATIONS

The following are the basic specifications for the PEL-3000 series. For detailed specifications, please see the user manual

Rating

Model	PEL-3021	PEL-3041	PEL-3111
Operating Voltage	1.5V~150V	1.5V~150V	1.5V~150V
Current	35A	70A	210A
Power	175W	350W	1050W

CC Mode Operating Range

Model	PEL-3021	PEL-3041	PEL-3111
H Range	0A~35A	0A~70A	0A~210A
M Range	0A~3.5A	0A~7A	0A~21A
L Range	0A~0.35A	0A~0.7A	0A~2.1A

CR Mode Operating Range

Model	PEL-3021	PEL-3041	PEL-3111
H Range	23.3336S ~400uS (42.857mΩ ~2.5kΩ)	46.6672S ~800uS (21.428mΩ ~1.25kΩ)	140.0016S ~2.4mS (7.1427mΩ ~416.6667Ω)
M Range	2.33336S ~40uS (428.566mΩ ~25kΩ)	4.6667S ~80uS (214.28mΩ ~12.5kΩ)	14.0001S ~242.4uS (71.427mΩ ~4.16667kΩ)
L Range	0.233336S ~4uS (4.28566Ω ~250kΩ)	0.46667S ~8uS (2.1428Ω ~125kΩ)	1.40001S ~24.24uS (714.27mΩ ~41.6667kΩ)

CV Mode Operating Range

Model	PEL-3021	PEL-3041	PEL-3111
H Range	1.5V~150V	1.5V~150V	1.5V~150V
M Range	1.5V~15V	1.5V~15V	1.5V~15V

CP Mode Operating Range

Model	PEL-3021	PEL-3041	PEL-3111
H Range	17.5W ~175W	35W~350W	105W ~1050W
M Range	1.75W ~17.5W	3.5W~35W	10.5W ~105W
L Range	0.175W~1.75W	0.35W~3.5W	1.05W ~10.5W

Slew Rate CC Mode Setting Range

Model	PEL-3021	PEL-3041	PEL-3111
H Range	2.5mA/us ~2.5A/us	5mA/us ~5A/us	16mA/us ~16A/us
M Range	250uA/us ~250mA/us	500uA/us ~500mA/us	1.6mA/us ~1.6A/us
L Range	25uA/us ~25mA/us	50uA/us ~50mA/us	160uA/us ~160mA/us

Slew Rate CR Mode Setting Range

Model	PEL-3021	PEL-3041	PEL-3111
H Range	250uA/us ~250mA/us	500uA/us ~500mA/us	1.6mA/us ~1.6A/us
M Range	25uA/us ~25mA/us	50uA/us ~50mA/us	160uA/us ~160mA/us
L Range	2.5uA/us ~2.5mA/us	5uA/us ~5mA/us	16uA/us ~16mA/us

EC Declaration of Conformity

We

GOOD WILL INSTRUMENT CO., LTD.

No.7-1, Jhongsing Rd., Tucheng Dist., New Taipei City 236,
Taiwan

GOOD WILL INSTRUMENT (SUZHOU) CO., LTD.

No. 69, Lushan Road, Suzhou New District Jiangsu, China

declares that the below mentioned product

PEL-3021, PEL-3041, PEL-3111, PEL-3020, PEL-3070

is herewith confirmed to comply with the requirements set out in the Council Directive on the Approximation of the Laws of the Member States relating to the Low Voltage Directive (2006/95/EC) and Electromagnetic Compatibility (2004/108/EC). For the evaluation regarding the Electromagnetic Compatibility and Low Voltage Equipment Directive, the following standards were applied:

◎ EMC

EN 61326-1 : EN 61326-2-1: EN 61326-2-2:	Electrical equipment for measurement, control and laboratory use — EMC requirements (2006)	
Conducted and Radiated Emissions EN 55011: 2009+A1: 2010	Electrostatic Discharge EN 61000-4-2: 2009	
Current Harmonic EN 61000-3-2: 2006+A1: 2009+A2: 2009	Radiated Immunity EN 61000-4-3: 2006+A1: 2008+A2 :2010	
Voltage Fluctuation EN 61000-3-3: 2008	Electrical Fast Transients EN 61000-4-4: 2004+A1: 2010	
-----	Surge Immunity EN 61000-4-5: 2006	
-----	Conducted Susceptibility EN 61000-4-6: 2009	
-----	Power Frequency Magnetic Field EN 61000-4-8: 2010	
-----	Voltage Dips/ Interrupts EN 61000-4-11: 2004	

◎ Safety

Low Voltage Equipment Directive 2006/95/EC
Safety Requirements: EN 61010-1: 2010; EN 61010-2-030: 2010

