

Specifications

GasAlertMicroClip XL Specifications

The instrument must satisfy the following:

Physical Specifications	
Size (h x w x d)	Physical size shall be no larger than 4.4 x 2.4 x 1.2 in. / 11.3 x 6.0 x 3.2 cm
Weight	Weight of instrument shall be no more than 6.7 oz. / 190 g, including alligator clip.
Case Material	Case material shall be rugged, impact-resistant, non-corrosive composite material that will prevent spark generation and shall be available in highly visible safety yellow or black.
Integral Boot	The instrument shall be equipped with a built-in concussion-proof boot.
Vibrator Alarm	The detector shall be equipped with an internal vibrator alarm for high noise areas.
Handling	Unit shall easily attach to pocket or belt.
Carrying Attachments	Unit shall be equipped with a high-tension stainless steel alligator clip.
Accessories	The following accessories must be available: • A manual aspirator pump with sampling hose or probe • A vehicle charger • Confined space entry kits in a hard-sided carrying case • Automatic Test and Calibration Station that shall: - o Bump test and calibrate instruments in the field without requiring use of a desktop PC or laptop - o Be fully portable and available in a variety of portable kit options - o Provide simultaneous management of up to 6 instrument docking modules - o Include battery and 110-240 VAC line power options - o Include software for downloading, evaluating and archiving monitoring results
User Interfaces	
Display Type	Large alphanumeric LCD (Liquid Crystal Display) type display with simultaneous and continuous readouts for each gas monitored in ppm, %LEL and % v/v values, as applicable, showing the real-time gas concentrations present. Digits not to be less than 0.260 in. / 6.50 mm high.
Display Symbols (Icons)	Gas alarm icons will clearly identify alarm type, gas hazard and alarm level encountered, low battery condition, instrument status, safe display, stealth, and calibration steps.
Keypad/Switches	One-button operation must: activate the detector; run the self-test; display alarm setpoints; display peak gas exposures and clears them; calibrate and zero. There shall be no requirement to access hidden switches for any instrument operation.

Specifications

Monitoring Capability			
Configurations	The gas detector must be available in 1, 2, 3, or 4 gas models that continuously and simultaneously monitor for oxygen, combustible/methane gases, hydrogen sulfide, carbon monoxide as applicable.		
Gases Detected and Measuring Specifications	Gases	Standard Measuring Range	In Increments of:
	Hydrogen Sulfide	0-100 ppm H ₂ S	1 ppm
	Carbon Monoxide	0-500 ppm CO	1 ppm
	Oxygen	0-30.0 % O ₂	0.1 %
	Methane	0-5% v/v methane	0.1% methane
	Combustible gases	0-100% LEL	1.0% LEL
Sensor Type	Sensors must be plug-in.		
Longer Life	The catalytic (combustible/methane) sensor should not be affected by common sulfur compounds, such as H ₂ S.		
Instrument Power			
Battery Type and Run Time	Lithium-polymer rechargeable battery pack. The run time of a new instrument shall not be less than 18 hours at room temperature. The run time shall not be less than 12 hours at any time over 2 years of use in a temperature range between -4°F to +122°F (-20°C to +50°C)		
Rechargeable Choices	A Vac (110 to 240) line charger, 12 Vdc (vehicle) charger and 12 Vdc / 24 Vdc cable must be available.		
Charge Time	Typical charge time for rechargeable batteries shall not be more than 6 hours per battery.		
Environmental			
Temperature Range	Normal operation: -4 to +122 °F / -20 to +50 °C		
Humidity	0-95% RH (non-condensing) continuous		
Ingress Protection	IP 68		
Calibration			
Automatic Calibration	Calibration must be fully automatic with Auto Zero and Auto Span functions. Instrument must advise as each automatic function takes place and when to apply gas. Calibration span levels must be “User Settable”.		
Calibration Diagnostics	Equipped with calibration diagnostics protection that ensures a valid calibration, the detector must check the ambient air and the calibration gas. If either does not meet expected values, the detector will refuse calibration and automatically exit the procedure, retaining all previous values.		

Specifications

Basic Operational Features

Instrument Activation	One-button activation. ON function must: • Test the battery and advise condition • Display the current alarm set points • Provide a full function self-test of sensor integrity, circuitry integrity and alarm activation • Automatically calibrate the oxygen sensor • Automatically zero H₂S, CO and LEL sensors • Advise when the next calibration is due in days, or advise the number of days overdue
Inadvertent Shut-off	Unit must not be able to be turned OFF accidentally and will incorporate a timed off function that provides audible/visual/vibration OFF indications.
Peak Exposures	Records peak (maximum) exposure to gases encountered and must display the reading on demand.
Accumulated Exposures	Records both TWA (time-weighted average) to toxic gases based on an 8-hour workday and STEL (short-term exposure limit) to toxic gases based on a 5-15 minute user-selectable work period. (STEL period is adjustable.) Must display readings on demand.
Confidence Beep	Confidence beep must be a user selectable option.
IntelliFlash (Compliance Flash)	IntelliFlash verifies operation and compliance to both the user and supervisor.
Instrument Status Advise	The detector must constantly analyze and test its own operational status and provide alarm advice of any malfunction.
Backlight	Backlight automatically illuminates the display in all alarm conditions (auto with time-out) and can be reactivated on demand with a button press.
Adjustable Options	The unit shall be equipped with user options for the following functions: • Adjust STEL period (5-15 minutes in 1 minute intervals) • Set calibration span levels and due date • Toggle ON/OFF: confidence beep, oxygen auto-calibration at start-up, Auto-zero LEL, CO and H₂S on start-up, latching alarms, "SAFE" display function (does not display gas concentrations unless readings change), stealth mode for silent operations, Force calibration when calibration overdue, Calibration lockout: calibration must be invoked via IR, mandatory bump test option, bump test reminder • Multi-language LCD selection in English, French, German, Portuguese and Spanish • Select combustible gases measurement: 0-100% LEL (Lower Explosive Limit) or Methane gas 0-5.0% v/v • Alarm setpoint disable (set to zero for off). • Calibration due date (1 to 365 days, or set to zero for off) • User-settable confidence beep • Individual sensor enable/disable.
Auto Zero Protection	Must be able to activate Auto Zero at any time. To ensure a valid Auto Zero, the instrument will first test the ambient air for background interfering gas before proceeding with the Auto Zero function.

Specifications

Instrument Alarms	
Alarms and Type	Simultaneous visual display alarms, audible alarms and vibrator alarms must warn in the event of a gas alarm condition, sensor fault or instrument status alarm.
Number of Gas Alarms	Must be equipped with four (4) user-settable alarm levels, including instant low, instant high, TWA (Time-weighted average) and STEL (short-term exposure limit). Also, will include an OL (over limit) alarm.
Visual Alarms	Instrument must be equipped with at least three flashing alarm bars easily visible from multiple angles. The backlight must light in all alarm conditions. LCD must provide positive clear alarm advice as to which alarm level and type has been exceeded (instant LOW, instant HIGH, TWA, STEL and/or over limit).
Audible Alarm	Instrument must be equipped with one variable pulsed audible alarm that shall be rated at 95+ dB.
Vibrator Alarm	Instrument must be equipped with an internal vibrator alarm for high noise areas.
Alarm Set points	Alarm set points must be adjustable. The current alarm set points shall be displayed each time on start-up.
Over Limit Exposure Protection	For exposures above the instrument's measuring range for the combustible sensor, the instrument will provide an over limit latching alarm that must be acknowledged.
Low Battery Warning	Must be equipped with low battery audible and visual alarms.
Datalogging	
Recorded	All events and occurrences.
Data Storage Period	With a default 15-second sampling rate interval the storage capacity will be 16 hours.
Compatibility	Data must be compatible for use with standard office programs, such as Excel and Access.
Event Logging	
Recorded	Must record the last 10 alarm events encountered. Information shall include: gas monitored, alarm level (in ppm or %) encountered, alarm duration in minutes and seconds, time elapsed since each alarm event occurred, life remaining, and cumulative alarm time.

Specifications

Certifications and Approvals	
Intrinsic Safety	Instrument must be certified by the following standards: • Approved by CSA to both U.S. and Canadian Standards CAN/CSA C22.2 No. 157 and C22.2 152 ANSI/UL - 913 and ANSI/ISA - 12.13.01 Part 1 CSA Class I, Division 1, Group A, B, C, and D • ATEX: CE 0539 g II 1 G Ex ia IIC T4 Ga Sira 13ATEX2330 EN 60079-0, EN 60079-11, and EN 60079-26 • IECEX: Ex ia IIC T4 Ga IECEX CSA 05.0015 IEC 60079-0, IEC 60079-11, IEC 60079-26 • KTL 12-KB4BO-0659X • EAC Certificate: RU C-GB.ГБ05.B.01115
Manufacturing Approval	The instrument must be certified compliant with ISO 9001 provisions.
RFI/EMI Protection	RFI/EMI protection must comply with EMC directive 2004/108/EC
Warranty	
Warranty	Two-year full, non-prorated warranty including instrument, sensors and battery. <i>Battery is guaranteed to have 12 hour runtime during warranty period under normal operating temperature of 4°F / -20°C to 122°F / 50°C.</i>