






THIS IS ISSUED IN STRICT CONFIDENCE THAT IT IS NOT USED AS A BASIS FOR MANUFACTURE OR SALE AND THAT IT IS NOT COPIED, REPRINTED OR DISCLOSED TO A THIRD PARTY EITHER WHOLLY OR IN PART WITHOUT THE PRIOR WRITTEN CONSENT OF CAL TEST ELECTRONICS, INC.

- NOTES:
- MATERIAL:
INSULATOR: PP-704
JACK: BRASS, Ni PLATED
 - RATING:
CURRENT: 30 A MAX.
RESISTANCE: $\leq 5\text{ m}\Omega$
 - RoHS 2 COMPLIANT



ISOMETRIC VIEW

THIRD ANGLE PROJECTION



REFERENCE

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN MILLIMETERS TOLERANCES ARE:
X.X = ± 0.5 X.X* = $\pm 0.5^{\circ}$
X.XX = ± 0.25
X.XXX = ± 0.125
DO NOT SCALE DRAWING

MATERIAL
SEE NOTES

TREATMENT
N/A

FINISH
SEE NOTES

PROJECT NO.

APPROVALS

DATE

DRAWN GSG

10/11/2021

CHECKER B. HANSEN

10/11/2021

ENGR. B. HANSEN

6/1/2021

DESIGN ACTIVITY

METRIC



22820 Savi Ranch Prky.
Yorba Linda, CA 92887
USA

TITLE

TERMINAL BLOCK ADAPTER

SIZE

B

CAGE CODE

43F45

DWG. NO.

CT4425

REV.

A

SCALE 2:1

SHEET 1 OF 1

REVISIONS

REV.

DESCRIPTION

DATE

APPROVED

CT4425-2

RED

-

CT4425-0

BLACK

-

PART NUMBER

CONFIGURATION

NOTES / REFERENCE

PART NUMBER TABLE