

HIOKI

9726 • 9727 • 9728 • 9729

PC CARD

Notes on PC card Usage

August 2005 Edition 1 Printed in Japan
9726A980-00 05-08H

HIOKI E. E. CORPORATION

HEAD OFFICE

81 Koizumi, Ueda, Nagano 386-1192, Japan
TEL +81-268-28-0562 / FAX +81-268-28-0568
E-mail: os-com@hioki.co.jp URL <http://www.hioki.co.jp/>

HIOKI USA CORPORATION

6 Corporate Drive, Cranbury, NJ 08512, USA
TEL +1-609-409-9109 / FAX +1-609-409-9108

Firmware upgrade requirements

To use this PC card with the following HIOKI models, the instrument(s) must be upgraded to the corresponding firmware versions as indicated below.

• 3196 POWER QUALITY ANALYZER:

Upgrade to Ver.1.40 or later

Please submit your request for the version upgrade firmware through our website at (<http://www.hioki.co.jp>), or contact your local HIOKI distributor.

• 8730-10 or 8731-10 WAVE COMPARATOR:

Upgrade to Ver.1.30 or later

Please contact your local HIOKI distributor directly for upgrade information.

The version number for each model will appear on the instrument's display when the power is turned on.

Instruments NOT compatible with Model 9729 PC CARD 1G

- 3169-20 and 3169-21 CLAMP ON POWER HITESTER
- 3196 POWER QUALITY ANALYZER

Please use Models 9726 PC CARD 128M, 9727 PC CARD 256M or 9728 PC CARD 512M with the above models.