DPM 942

No.:695 - Page 1 of 2

A loop powered indicator giving an accurate, zero corrected indication of 4-20mA loop current. Calibration is by two 20-turn cermet potentiometers which allow sensitive adjustment of the instrument. The meter is housed in a robust carrier, which can be bolted in place or panel mounted using the low profile bezel and clips provided.

- (19mm (0.75") Digit Height
- **()** Programmable Decimal Points
- Low Volt Drop
- **U** Simple Screw Terminal Connections



CUSTOM SCALING

For alternative scaling to the standard 0 - 1000, cut Link 8 and short Link 7. Add two resistors R12 and R13 which will set the basic scale of the meter. The highest practical value for the R12 and R13 series combination is 12R. To select your desired scaling, use the following formula: R12 + R13 = Full Scale reading at 20mA

160 Re-calibrate as necessary.



Stock NumberStandard MeterDPM 942				
Specification	Min.	Тур.	Max.	Unit
Accuracy (overall error) *	0.05		0.1	%(±1 count)
Linearity			±1	count
Sample rate		3		samples/sec
Operating temperature range	0		50	°C
Temperature stability		100		ppm/°C
Loop Volt Drop	3.9	4.3	4.8	V

*To ensure maximum accuracy, re-calibrate periodically.

CONNECTOR SOURCING GUIDE

METHOD Screw Terminals - No Connector Required





LASCAR ELECTRONICS LTD. MODULE HOUSE, WHITEPARISH, WILTSHIRE SP5 2SJ UK TEL: +44 (0)1794 884567 FAX: +44 (0)1794 884616 E-MAIL: sales@lascar.co.uk

LASCAR ELECTRONICS INC. 4258 WEST 12th STREET, ERIE, PA 16505 USA TEL: +1 (814) 835 0621 FAX: +1 (814) 838 8141 E-mail: us-sales@lascarelectronics.com

LASCAR ELECTRONICS (HK) LTD. 8th FLOOR, CHINA AEROSPACE CENTRE, 143 HOI BUN ROAD, KWUN TONG, KOWLOON, HONG KONG TEL: +852 2389 6502 FAX: +852 2389 6535 E-mail: saleshk@lascar.com.hk

TERMINAL FUNCTIONS

- 1. I+ Positive current input.
- 2. I-Negative current input.
- 3. DP3 1.000
- 4. DP2 10.00 Decimal point inputs.
- 5. DP1 100.0
- 6. DS Decimal Point select. Connect to required DP input to display decimal point.

CALIBRATION

The meter is supplied calibrated to read 000 for 4mA loop current and 1000 for 20mA.

To re-calibrate:

- 1. Apply 4mA to I+/I- and adjust 'OFFSET' to read 000.
- 2. Apply 20mA and adjust 'SPAN' to read 1000.
- 3. Repeat steps 1 and 2 until there is no more adjustment of 'SPAN' and 'OFFSET' required to give desired readings.

SAFETY

To comply with the Low Voltage Directive (LVD 93/68/EEC), input voltages to the module's pins must not exceed 60Vdc. If voltages to the measuring inputs do exceed 60Vdc, then fit scaling resistors externally to the module. The user must ensure that the incorporation of the DPM into the user's equipment conforms to the relevant sections of BS EN 61010 (Safety Requirements for Electrical Equipment for Measuring, Control and Laboratory Use)



